1. **بالنسبة للإفادة المصرفية : يجب ان يكون المبلغ اكبر من المبلغ النهائي في عرض الأسعار او ان يتعهد المصرف بتأمين السيولة اللازمة في حال رسو التلزيم**
2. **ان اثبات الخبرة يجب ان يتضمن الأسس الذي تم على أساسه التلزيم (عقود او أوامر الشراء) ويجب توافر عقدين او امرين شراء على الأقل بالإضافة الى إفادات حسن التنفيذ او محاضر الاستلام النهائية**

3 – Transducer Frequency Range

The frequency bandwidth of each offered transducer shall be suitable for the clinical applications specified in the Technical Specifications.

4 – Additional Spare Probes

Please refer to the scoring breakdown under criterion R3. Each additional probe must have a frequency bandwidth suitable for the clinical applications specified for that probe type.

5 – MoPH-Approved Reference Countries

Please refer to the scoring breakdown under criterion R1. The points for this criterion are based on the location of the manufacturer's headquarters, as assessed against the list of reference countries approved by the Ministry of Public Health. Installation references are not used to determine these points.

6 – Eligibility of Ultrasound Systems Manufactured in China

Ultrasound systems manufactured in China are eligible for consideration, provided they meet all mandatory requirements of the bidding documents. The location of the manufacturer's headquarters is considered solely for scoring under criterion R1

7 – Intraoral Dental X-ray System (J1)
Either a portable or a wall-mounted intraoral dental X-ray system is acceptable,

8 – Electronic Apex Locator (S4)
A standalone electronic apex locator or an apex locator integrated into the offered endodontic motor is acceptable

9 – Complete Implantology System (S5)
A built-in or standalone implantology system is acceptable

10 – Air Polishing Unit (S6)
A built-in, standalone, or connectable air polishing unit is acceptable

11- ان السعر المقدم يجب ان يتضمن كامل المصاريف و الضرائب

12- ان التسليم يكون في مراكز الرعاية الصحية (في الاقضية والمحافظات)

13-ان مهلة التسليم تتضمن التسليم والتركيب والتشغيل